

# Physical-Technical Testing Institute Ostrava – Radvanice

### **Supplementary Type Examination Certificate No.10**

about verification of workplace atmospheres – Electrical apparatus used for the direct detection and direct concentration measurement of toxic gases and vapours

Type Examination Certificate number:

### FTZÚ 09 E 0026

Product: ALTAIR 5X Multigas Detector

Manufacturer: MSA - The Safety Company

Address: 1000 Cranberry Woods Drive, Cranberry Twp, PA 16066, USA

This supplementary certificate extends Certificate No. FTZÚ 09 E 0026 to apply to products designed and constructed in accordance with the specification set out in the Schedule of the said certificate but having any variations specified in the Schedule attached to this certificate and the documents therein referred to.

The Physical-Technical Testing Institute, Certification Body No.3051, accredited by ČIA o.p.s, Prague according to EN ISO/IEC 17065:2012, certifies that this product has been found to comply with requirements of the following standard:

EN 45544-1:2015, EN 45544-3:2015, EN 50104:2010, EN 50271:2010

Manufacturer listed in the certificate is responsible for the product conformity assurance in accordance with its specification (documentation) listed in this certificate and for successful performance of all specified routine tests and verification.

This supplementary certificate relates only to the verification of workplace atmospheres – Electrical apparatus used for the direct detection and direct concentration measurement of toxic gases and vapours. Further requirements can be applied to the manufacturing process and supply of this product. These are not covered by this certificate.

This certificate is valid till: 04.04.2024

Responsible person:

Dipl. Ing. Lukáš Martinák Head of certification body



Date of issue: 29.05.2019

Page: 1/2



# Physical-Technical Testing Institute Ostrava – Radvanice

#### Schedule

## Supplementary Type Examination Certificate No. 10 to FTZÚ 09 E 0026

Description of product changes:

The firmware has been updated to version v2.19.1. The assessment in scope of the standard EN 50271:2010 has been done.

The assessment according to other standards EN 45544-1:2015, EN 45544-3:2015 and EN 50104:2010 has not been affected.

#### Recapitulation of parameters:

Performance tests have been carried out for:

- Gas H<sub>2</sub>S in range 0-200ppm by electrochemical sensor. The TWA function is not supported.
- Gas O2 in range 0-25% (V/V) by electrochemical sensor. The warm-up time is up to 180 sec.

Test report No.: 09E/0026/10

List of documentation:

Number:

Rev.

Title:

SK 3098-1184

9

FTZU approvals, ALTAIR5X / ALTAIR5XIR

Responsible person:

Dipl. Ing. Lukáš Martinák Head of certification body



Date of issue: 29.05.2019

Page: 2/2